CAN CEREBRAL OXYGENATION REFLECT BLOOD LOSS? - A CASE SERIES IN TOTAL HIP REPLACEMENT SURGERY

Joana Tavares¹, Joana Teixeira², Pedro Amorim¹, Manuela Araújo¹
¹ CH Porto, ² CH Setúbal

Background: Cerebral oximetry estimates the oxygenation of cerebral cortex by a non-invasive transcutaneous measurement using near-infrared spectroscopy (NIRS). Although it has been used to detect cerebral ischemia, its utility to monitor blood loss is yet to be confirmed ^{2,3}.

Goal: to study if cerebral oxygenation (cerebral rSO₂) changes induced by the normoxing challenges (↓FiO₂ from 50% to 21% in ventilated patients) can reflect the amount of blood lost.

Materials and methods:

- ASA 1-2 patients undergoing general anesthesia for total hip replacement surgery were prospectively studied.
- Patients were monitored by ASA standards.
- Depth of anesthesia was monitored by BIS and cerebral oxygenation (rSO₂) by INVOS.
- Every 30 minutes a normoxic challenge was performed (decrease of FiO₂ from 50% to 21%) and an arterial blood sample was obtained to measure Hgb and blood gases. Heart rate (HR), mean arterial pressure (MAP), SpO₂, EtCO₂, BIS, rSO₂(cerebral and thenar eminence) and estimated blood loss were also recorded every 30 minutes.
- Statistic analysis used Friedmann test, Spearman correlation and linear regression (SPSS Statistic 23).

Results and discussion:

10 patients were studied but only 9 had at least three normoxic challenges (total of 29). Median values of HR, MAP, BIS, EtCO₂, SpO₂, rSO₂(cerebral and thenar) are presented in Table 1. There were no correlations between HR, MAP, EtCO₂ and BIS and the challenges neither with Hgb and blood gases.

We found a correlation between consecutive challenges and the fall of cerebral rSO₂ (p=0,01) and a trend towards an increase of that fall over time (Graph 1). SpO₂ showed no correlations with other variables. Estimated blood loss increased with the length of the surgery (p<0,001) (Graph2) and there was a correlation with cerebral the rSO₂ variation (p=0,002) (Graph 3). Cerebral rSO₂ with 50% FiO₂ did not change during surgery. Hemoglobin value did not correlate with any parameter.

	Fi O ₂ 50%	1st challenge	2nd challenge	3rd challenge
HR (bpm)	68 (51-85)	64 (56-80)	69 (54-81)	71 (51-76)
MAP (mmHg)	84 (60-112)	79 (59-97)	82 (66-93)	76 (66-88)
BIS	39 (28-49)	41 (34-58)	42 (36-63)	52 (30-61)
EtCO ₂	32 (29-39)	34 (29-39)	34 (30-39)	34 (31-41)
SpO ₂	98 (96-99)	98 (95-99)	98 (96-100)	98 (97-100)
rSO ₂ thenar (%)	65 (43-85)	66 (48-75)	68 (42-74)	65 (55-78)
rSO ₂ cerebral (%)	77 (57-85)	72 (59-81)	74 (55-84)	74 (54-84)

Table 1: median (maximum-minimum) values of HR, MAP, EtCO₂, SpO₂ and rSO₂ in the three challenges and with FiO₂ 50%.

Cerebral rSO₂ variation Δ_{50-21} (%) in consecutive challenges

Graphic 1: variation (in percentage) of cerebral oxygenation values, before and after FiO₂ decrease from 50 to 21%, between consecutive challenges.

Blood loss (ml) in consecutive challenges

Graphic 2: variation of blood loss (in milliliters) between the three consecutive challenges.

Cerebral rSO₂ variation Δ_{50-21} (%) and blood loss (ml)

Graphic 3: relationship between cerebral oxygenation values (in percentage), before and after FiO₂ decrease from 50 to 21%, and the amount of blood loss (in milliliters).

Conclusions: Performing normoxic challenges provided additional clinically useful information as it showed the cerebral perfusion state without the contaminant factor of the high FiO₂. We found a correlation between the amount of blood loss and the decrease of cerebral rSO₂, although there was no decrease in hemoglobin measurements. This lack of correlation with Hgb can be explained by the fact that hipoxia is caused by a mismatch between oxygen delivery and demand in tissues, not always translated into immediate hemoglobin values. So, we hypothesize that it was caused by blood loss.

References: ¹Ferari M, Principles, techniques and limitations of near-infrared spectroscopy, Can J Appl Physiol. 2004; 29, ²Murkin JM, Near-infrared spectroscopy as an index of brain and tissue oxygenation, BJA, 2009;103, ³Torella F, Cerebral and peripheral oxygen saturation during red cell transfusion, J Surgical Research,2003;110