Assessment of the quality of chest compressions in healthcare professionals

Artem Kuzovlev^{1,2}, Viktor Moroz^{1,2}, Arkady Goloubev^{1,2},
Sergey Abdusalamov², Kirill Kuzmichev², Evgeny Tishkov³

¹ Federal Research and Clinical Center of Intensive Care Medicine and Rehabilitology, Moscow, Russia

² Russian National Resuscitation Council, Moscow, Russia

³ A.I. Evdokimov Moscow medical dental university, Moscow, Russia

- **Immediate and high-quality chest compressions (CC)** make the background of in-hospital cardiopulmonary resuscitation.
- **None of the studies** on feedback or prompt devices has demonstrated improved survival to discharge with feedback [ERC guidelines 2015].
- The use of CPR feedback or prompt devices during CPR should only be considered **as part of a broader system of care** that should include comprehensive CPR quality improvement initiatives, rather than as an isolated intervention [ERC guidelines 2015].

The aim of the study was to assess the quality of chest compressions in healthcare providers in the multidisciplinary hospital.

- The study was performed in a multifield hospital in Moscow, in 2016-2017.
- 42 healthcare professional were enrolled.
- Chest compressions were assessed in 2 steps by means of **CPR-challenge** tool: 2 mins of CC without feedback then 2 mins of CC with feedback on rate, depth and recoil
- **The ERC Guidelines for Resuscitation 2015** were used as reference criteria for CC quality.
- The analysis was performed using the ZOLL RescueNet Code ReviewR software.
- Statistical analysis was performed using the Statistica 7.0 software (Mann-Whitney *U*-test). The data were presented as a mean, median $\pm$ 25—75 percentiles (25—75 IQR), minimum and maximum values. The difference was considered significant at $P < 0.05$.

- **Without feedback device** the percentage of CC in target (with correct rate, depth and recoil) was $5,4 \pm 18,3\%$ in 86% of healthcare providers.
- **With feedback device** the percentage of CC in target was $68,3\% \pm 27,4\%$ ($p = 0,000$).

Chest compressions quality without controller

Chest compressions quality with controller

**Healthcare providers have low background CC skills which improve when feedback devices are used.
Regular simulation training and retraining is required.**