Development of the first commercial real-time PCR assay for the detection of Mucorales species (MucorGenius)

Giel Gaajetaan¹, Dennis van Tegelen¹, Tim Kampermann¹, Gijs Dingemans¹

¹ PathoNostics B.V., Maastricht, The Netherlands

Introduction

Mucormycosis is a rare invasive fungal infection caused by Mucorales species and generally affects severely immunocompromised individuals. Mucormycosis is associated with a high mortality rate and prevalence is underestimated due to the limited diagnostic options. Correct and fast identification of mucorales infection is essential to start appropriate antifungal therapy.

Due to this limitation in current diagnostics, PathoNostics developed a real-time PCR for the detection of the clinical most relevant Mucorales species. The goal of this study was to determine the analytical specificity, inclusivity and sensitivity. The performance of the MucorGenius PCR was also examined on a set of ring trial samples.

Materials & Methods

The real-time PCR assay (MucorGenius, PathoNostics) enables detection of the 28S multicopy gene in Mucorales species together with an internal control.

In order to determine the analytical sensitivity, serial DNA dilutions of 7 reference strains including *R. oryzae*, *M. racemosus*, *L. corymbifera*, *R. microsporus*, *C. bertholletiae*, *R. pusillus* and *R. miehei* were quantified using droplet digital PCR (ddPCR).

Quality assessment was performed using the ring trial of the Fungal PCR Initiative (FPCRI) for Mucorales species.

Results

The MucorGenius PCR was able to specifically detect all tested Mucorales species. No cross-reactivity was observed with other fungi, yeast or bacterial strains commonly present in the respiratory tract (data not shown).

The analytical sensitivity was determined as the highest dilution of Mucorales DNA which could be detected with the MucorGenius and is quantified with ddPCR (figure 1). All MucorGenius targets were detected at 1-2 copies/μl of extracted DNA which corresponds with a Ct-value of 35/36 on the LC480 II real-time PCR (table 1).

The MucorGenius PCR identified Mucorales in 5 out of 6 FPCRI samples, which was in agreement with the FPCRI data (table 2).

Conclusion

- ❖ The MucorGenius real-time PCR from PathoNostics can specifically detect the clinically most relevant Mucorales species in low concentrations.
- ❖ The MucorGenius PCR had a 100% score with the FPCRI panel.
- ❖ Further validation on clinical samples is ongoing.

Sample nr.	Inoculated DNA	MucorGenius	Ct-value
S1	<i>R. pusillus</i>	+	34
S2	None	-	Negative
S3	<i>R. oryzae</i>	+	32
S4	<i>C. bertholletiae</i>	+	33
S5	<i>L. corymbifera</i>	+	34

Table 2. FPCRI Mucorales panel results.

Figure 1. Mucorales amplification curves of the highest, detectable DNA dilutions. Detection of *L. corymbifera*, *R. oryzae*, *M. racemosus*, *R. microsporus*, *C. bertholletiae*, *R. pusillus* and *R. miehei* DNA.

ddPCR results analytical sensitivity MucorGenius		
Target	Conc. (copies/μl)	Ct-value
<i>L. corymbifera</i>	1	35
<i>R. oryzae</i>	1	35
<i>M. racemosus</i>	1	35
<i>R. microsporus</i>	1	36
<i>C. bertholletiae</i>	2	35
<i>R. pusillus</i>	2	35
<i>R. miehei</i>	2	36

Table 1. Analytical sensitivity of the MucorGenius PCR for all tested Mucorales species. Ct-values were determined on the LC480 II PCR.