ADAPTING DEFIBRILATORS TO VISUALLY HANDICAPPED PEOPLE WITH A STICKER

Jorge-Soto C.^{a,b}, Martínez-Isasi S.^c, Varela-Casal C.^d, Abelairas-Gómez Ca.^e, Barcala-Furelos R.^{a,d}, Rodríguez-Núñez A.^{a,b}

^a CLINURSID Research Group, University of Santiago de Compostela. Spain

^b University of Santiago de Compostela. Spain

^c University of A Coruña. Spain

^d University of Vigo. Spain

PURPOSE: Visually impaired people respond to brief basic live support training programs in a similar way to no handicapped people. Thus, they are valid candidates to be rescuers in an out of hospital cardiac arrest. The aim of the present study is to propose a sticker which could ease the AED use by visually handicapped people.

MATERIAL AND METHODS

4 Blind People
+
31 Blindfolded

Naïve in BLS

- Very brief training
- All time in touch with training material (Laerdal manikin and AED)
- Participants/instructor ratio 5/1

1. **Practical test** in a simulated scenario.
2. **Survey** about difficulties perceived while using the AED and possible changes.

RESULTS

Main Difficulties Perceived

	BLIND PEOPLE	BLINDFOLDED
Finding the power on button	--	6/35
Finding electrodes	3/4	5/31
Plug in electrodes	--	13/31
Finding shock button	3/4	5/31

The Most Proposed Changes

CONCLUSIONS

Using a defibrillator without a sense of sight, participants found that signalling of different defibrillator elements could not be perceived. Thus, we propose an adhesive template that includes signage to adapt defibrillators to people with visual impairments.