Department of Anaesthesiology and Intensive Care Medicine | Head: Univ.-Prof. Dr. Bernd W. Böttiger

Accuracy and precision of the non-invasive Zero-Heat-Flux thermometry in conditions of cardiac arrest

Tobias Neumann, Simon-Richard Finke, Alexandra C. Maul, Maja Henninger, Ben Höpfner, Sebastian Lemke, Thorsten Annecke, Daniel C. Schroeder*

* Center for Experimental Medicine, University Hospital of Cologne

Purpose of the study

Precise measurement of core body temperature is obligatory for targeted temperature management (TTM). Temperature measuring devices may be unprecise, accompanied by side-effects and/or application could be labour-intensive.

Objective was to evaluate accuracy and precision of the non-invasive Zero Heat Flux-thermometry (ZHF) to monitor core body temperature.

Materials and methods

With approval by the animal welfare authorities (LANUV NRW, reference 84-02.04.2017.A176), ventricular fibrillation was initiated in 19 pigs. Cardiopulmonary Resuscitation (CPR) was started after 5 minutes of untreated cardiac arrest (CA).

Core body temperature was continuously measured by ZHF (3M SpotOn™) and thermodilution-catheter (PULSION Medical Systems). Measurement-pairs were recorded

- a) prior to CA (n = 19) every 5 minutes,
- b) during CA (n = 17) every 30 seconds,
- c) during CPR (n = 17) every 60 seconds,
- d) within 60 minutes after return of spontaneous circulation (ROSC, n = 8) every 60 seconds.

Bias and Limits of Agreement (LoA) were calculated according to Bland and Altman.

Results

	Bias	Lower LoA	Upper LoA
a) Prior to CA	-0.21	-0.63	0.21
b) During CA	-0.13	-0.8	0.54
c) During CPR	-0.23	-1.19	0.74
d) During ROSC	-0.05	-0.84	0.73

Reference List

- 1) Pei L, Huang Y, Mao G, Sessler DI. Anaesth Analg 126 (3): 833-838, 2018. <https://doi.org/10.1213/ANE.0000000000002706>
- 2) Guschlbauer M, Maul AC, Yan X, Herff H, Annecke T, Sterner-Kock A, Böttiger BW, Schroeder DC. PLoS One 11(3): e0150759, 2016. <https://doi.org/10.1371/journal.pone.0150759>
- 3) Leary M, Cinousis MJ, Mikkelsen ME, Gaieski DF, Abella BS, Fuchs BD. Resuscitation 85 (2):244-247, 2014. <https://doi.org/10.1016/j.resuscitation.2013.10.027>.

By colouring the measurement-pairs post ROSC along a time-dependent scale it becomes visible that especially precision increases over time.

To further analyse this time-dependency in the post ROSC period we separated this period into two intervals:

- e) ROSC 0 to 30 minutes,
- f) ROSC 31 to 60 minutes.

	Bias	Lower LoA	Upper LoA
e) ROSC 0 - 30 min	-0.10	-1.01	0.81
f) ROSC 31 - 60 min	-0.03	-0.62	0.57

Conclusions

Prior to CA, both accuracy and precision of ZHF ranged within $\pm 0.5^{\circ}\text{C}$, which is clinically acceptable and in line with preceding publications. [1, 2]

Based on clinical criteria, ZHF is unprecise during CA and CPR. Within the first 30 minutes post ROSC, precision of ZHF is insufficient to regulate TTM.

As precision increases time-dependently after achieved ROSC and complies to clinical standards 30 minutes later, TTM may be controlled by ZHF when initial stabilisation is usually finished. [3]