Comparison of the TOF-Cuff Monitor[®] with the TOF-Watch SX Monitor[®] An observational study

E. Sfeir Machado ^a, E. Dupuis-Lozeron ^b, G. Kely Barcelos ^a, MR Tramèr ^{a,c}, C. Czarnetzki ^{a,c}

a) Division of Anesthesiology, Department of Anesthesiology, Pharmacology & Intensive Care Medicine, Geneva University Hospitals, Geneva, Switzerland

b) Clinical Trials Centre & Division of Clinical Epidemiology, Department of Health and Community Medicine, University of Geneva & Geneva University Hospitals, Geneva, Switzerland

c) Faculty of Medicine, University of Geneva, Geneva, Switzerland

Background and Goal of Study

The TOF-Cuff[®] is a medical device that combines blood pressure cuff and neuromuscular monitoring.¹ It has never been validated using acceleromyography. We compared recovery parameters of a neuromuscular block between the TOF-Cuff[®] and the TOF-Watch SX[®]. The TOF-Watch SX device is well recognized in the research setting as a reference method.²

Materials and Methods

Nineteen patients, ASA I to II, aged 18-65 years, undergoing elective surgery, were included in this single centre study. The TOF-Cuff[®] was installed on one upper arm and the TOF-Watch SX[®] on the opposite wrist.

A single IV dose of rocuronium 0.6 mg kg⁻¹ was administered for intubation. Anaesthesia was maintained with a continuous propofol infusion. After calibration of both devices, continuous TOF stimulation was started. Primary outcome was the total recovery time (time in minutes from injection of rocuronium to a normalised TOF ratio of 90%). Bias and limits of agreement between the two monitors were calculated as proposed by Bland and Altman.

The Tof-Cuff boasts the working principle of *'Keep It stupid Simple'*.³ It simplifies the concept of measurement of neuromuscular block which is based on different sensitivity of muscle groups to neuromuscular blocking agents.

Results and Discussion

There were two monitoring failures in the TOF-Watch SX[®] group. In three patients, neuromuscular block had to be antagonised at the end of surgery. In 14 patients, the primary outcome could be analysed for both monitors. With the TOF-Cuff[®], total recovery time was 20 ± 23 min shorter compared with the TOF-Watch SX[®]. The difference was increasing with increasing total recovery time, suggesting that the bias of the TOF-Cuff[®] was not systematic.

Cuff
60%

Watch
8%

Cuff
100%

Watch
40%

Conclusion

The TOF-Cuff[®] overestimates recovery of a neuromuscular block compared with the TOF-Watch SX[®]. Patients who are extubated based on TOF-Cuff[®] values are at risk of residual neuromuscular block.

References

1. Rodiera J et al. The cuff method: a pilot study of a new method of monitoring neuromuscular function. Acta Anaesthesiol Scand. 2005.
2. Fuchs-Buder T et al. Good clinical research practice in pharmacodynamic studies of neuromuscular blocking agents II: The Stockholm revision. Acta Anaesthesiol Scand. 2007.
3. www.rgb-medical.com/en/special-product/tof-cuff-nmt-monitor (assessed 21.5.2018)

Acknowledgements

- Anel GmbH, Weyermannsstrasse 12, 3008 Bern www.anelmed.ch (For lending the Tof Cuff device)
- Hombourger JM for technical assistance

Funding: Institutional funds of the Division of Anesthesiology, Department of Anesthesiology, Pharmacology & Intensive Care, HUG.