Clinical outcomes following Cochlear™ BIA300 bone anchored hearing aid implantation in children

Jonathan Fussey, Emilie Harterink, Jaspreet Gill, Anne Child-Hymas,
Ann-louise McDermott

Birmingham Children's Hospital
Ear, Nose and Throat Surgery

Introduction

The Cochlear™ BIA300 system was launched in 2010 and used at The Birmingham Children's Hospital from 2011. Here we report the long-term outcomes of children implanted with the Cochlear™ BIA300 BAH system in our centre.

Methods

This was a retrospective 5-year longitudinal case records review of children who were implanted with the Cochlear™ BAI300 implant system at Birmingham Children's Hospital between December 2011 and December 2013. There were no exclusions.

Patients

- Mean age 8.7years, 58% male
- Significant co-morbidity in 37%
- Learning disability in 88%; mean BMI 19.1kg/m²

Surgery

- Three surgeons
- 18% single stage; 82% sequential
- Soft tissue reduction in 58%

Implants

- 58% 4mm; 42% 3mm according to bone thickness
- 63% 6mm abutment; 27% 9mm abutment

Results

52 children with a total of 78 implants were included. Mean follow-up was 43.5 months. There was a higher rate of skin overgrowth requiring a longer abutment in the group with learning disabilities (57%) compared to those without (31%, P=0.0380). A planned abutment removal was performed due to soft tissue reaction in 26% of cases. The fixture failure rate was 5%.

	Number of implants	(%)
Total	78	100
Total soft tissue complications	60	77
Requiring local treatment	48	62
Requiring systemic treatment	27	35
Requiring surgical revision	27	35
Requiring longer abutment	36	46
Requiring abutment removal	20	26

Conclusions

This study shows an improved implant survival rate with the wider Cochlear™ BI300 implant, but higher than expected rates of soft tissue complications with the BA300 abutment. Soft tissue complications are commoner in children undergoing BAH, and there are several factors that may contribute to the increased rates seen in this cohort, however the possibility that the implants themselves are at least partly responsible cannot be discounted.