Development of a needle for peripheral intravenous catheters to improve vein puncture using in vitro models

Tanabe H^{1,2}, Kawasaki M², Ueda T², Zushi Y², Murayama R^{1,3}, Abe M^{1,3}, Sanada H^{3,4}

¹ Department of Advanced Nursing Technology, Graduate School of Medicine, The University of Tokyo, Tokyo, Japan

² Research and Development section, TERUMO Corporation, Kanagawa, Japan

³ Global Nursing Research Center, Graduate School of Medicine, The University of Tokyo, Tokyo, Japan

⁴ Department of Gerontological Nursing/Wound Care Management, Graduate School of Medicine, The University of Tokyo, Tokyo, Japan

INTRODUCTION

Peripheral intravenous catheters (PIVCs) are commonly used in clinical practice. However, many PIVCs are unsuccessful with first attempt success rates as low as 50%.¹

One of the risk factor for this is advanced patient age. With the loss of subcutaneous fat and thinning of the skin in older adults, introduction of a needle into a vein can cause the vein to deform or move, making placement of a vascular access device challenging.²

OBJECTIVES

We have developed an ultra-sharp needle with a very thin tip for older patients. Bevel angle is same, but bevel face length is smaller than the regular bevel. We evaluated the needle using two in vitro models. The ultra-sharp needle in PIVC22G was compared with four types of PIVC22G by a resistance force test and tube puncture test.

CONCLUSIONS

The resistance forces at needle tip produced by the ultra-sharp needle were significantly low.

The needle may improve the rates of first puncture attempts in older patients. Further studies are required to demonstrate the effectiveness of this needle in clinical settings.

REFERENCES

- Idemoto, B.K., Rowbottom, J.R., Reynolds, J.D., Hickman Jr., R.L., 2014. The accucath intravenous catheter system with retractable coiled tip guidewire and conventional peripheral intravenous catheters: a prospective, randomized, controlled comparison. *J. Assoc. Vasc. Access* 19 (2), 94–102.
- Infusion Nurses Society. Policies and Procedures for Infusion Nursing of the Older Adult. *J Infus Nurs.* 77.

METHODS

Resistance force test measured the load when the needle tip and the catheter tip penetrated a polyethylene sheet.

Hypothesis: Resistance forces at the needle tip of the ultra-sharp needles are the lowest.

Tube puncture test evaluated whether the needle could puncture a polyvinyl chloride (PVC) tube (as a vein model).

Hypothesis: The ultra-sharp needles puncture the PVC tube at smaller angle both in the center and offset condition.

RESULTS

Resistance force test

N=10

Resistance forces of needle tip and catheter tip

(*p<0.05 vs. ultra-sharp; **p<0.01 vs. ultra-sharp, mean±SD)

Tube puncture test

Success rates of the tube puncture (center)

Angle	Ultra-sharp	A	B	C	D
15°	X 0/5	X 3/5	X 4/5	X 1/5	X 0/5
20°	✓ (5/5)	✓ (5/5)	✓ (5/5)	✓ (5/5)	✓ (5/5)
25°	✓ (5/5)	✓ (5/5)	✓ (5/5)	✓ (5/5)	✓ (5/5)
30°	✓ (5/5)	✓ (5/5)	✓ (5/5)	✓ (5/5)	✓ (5/5)

Success rates of the tube puncture (0.5mm offset)

Angle	Ultra-sharp	A	B	C	D
15°	X 0/5	X 0/5	X 0/5	X 0/5	X 0/5
20°	✓ (5/5)	X 2/5	X 0/5	✓ (5/5)	✓ (5/5)
25°	✓ (5/5)	X 3/5	X 2/5	✓ (5/5)	✓ (5/5)
30°	✓ (5/5)	X 4/5	X 3/5	✓ (5/5)	✓ (5/5)

DISCUSSION

- The ultra-sharp needles produced significantly lower resistance forces of needle tip in any other products.
- It was succeeded when the ultra-sharp needles punctured a PVC tube at smaller angle (20°).
- The ultra-sharp needle may prevent the vein to deform or move in older patients.

Type	Needle Geometric design	Bevel face length (mm)	Bevel angle (°)
A	Lancet	2.0	30
B		1.8	31
C	Back-Cut	1.7	40
D		1.8	46
Ultra-sharp		2.0	20

Conflict of interest

Tanabe H, Kawasaki M, Ueda T and Zushi Y are employed by TERUMO corporation. The other authors have been determined to have no conflict of interest. This research was a joint project with Terumo corporation, conducted with sponsorship from this organization.

Copyright © 2018 Tanabe H, e-mail: Hidenori_Tanabe@terumo.co.jp