Different methods to detect incomplete heparin reversal are not interchangeable in children after congenital cardiac surgery

Schenk S.¹, Nowak-Machen M.¹, Rosenberger P.¹, Magunia H.¹

¹University Hospital Tübingen, Department of Anaesthesiology and Intensive Care Medicine, Germany

Background

The use of patient-specific heparin concentrations is a recommended procedure in cardiac surgery and has been shown to lower bleeding complications [1, 2]. Incomplete heparin reversal can lead to excess bleeding after cardiopulmonary bypass (CPB). This study aimed to compare different parameters to detect incomplete heparin reversal in infants undergoing congenital cardiac surgery that were managed with individually calculated heparin and protamine doses.

Methods

After approval by the local IRB a total of 36 children undergoing congenital cardiac surgery (STS classes 1-5) between 02/2017 and 06/2017 were retrospectively enrolled. Besides the results of the HMS Plus system (Hepcon), preoperative and postoperative values of aPTT and ACT were analyzed and the ratio of the post- to preoperative (PPR) values of aPTT and ACT was calculated. The ratio of the time to clot initiation (TTC) of the intrinsic pathway assays (CK or INTEM) to heparinase based assays (CKH or HEPTM) by the TEG®6s and the ROTEM® delta system was also calculated. Ratios above 1 were considered as a sign for incomplete heparin neutralization.

Figure1: TEG6s® showing normal CKH-R time and prolonged CK-R time as an effect of heparin.

In a subset of 18 patients the TTC (CK-time ratio ROTEM® delta) detected 7/18 (39.9%) whereas the TTC (R-time ratio TEG®6s) detected 17/18 (94.4%). In this subset the level of agreement between the TEG®6s and ROTEM® delta based assays (complete reversal yes or no) revealed a phi coefficient of 0.193.

Discussion and Conclusion

The detection of incomplete heparin reversal is an important part of coagulation management after CPB because incomplete heparin reversal is a major reason for bleeding complications after cardiac surgery. Protamine itself can lead to a thrombocyte associated coagulation disorder. Our study shows that the studied tests are not interchangeable. The new TEG®6s system seems to be very "sensitive" in detecting incomplete heparin reversal. Further studies are warranted to analyze if this truly reflects incomplete heparin reversal. Test results should always be correlated with clinical signs of bleeding prior to therapy.

References

- [1] Gautam NK, et al. (2013). Paediatr Anaesth.;23:233-41
- [2] Ferraris VA, et al. (2011). Ann Thorac Surg.;91:944-82

	Median	Interquartile Range	Min.-Maximum
Age (d)	196	39,5 – 573	4 – 1704
Duration of CPB (min)	93	63 – 138	18 – 302
Dose of Heparin (I.E./kg)	455	348 – 646	-
Dose of Protamine (I.E./kg)	428	311 – 599	-
pre-aPTT (s)	28	25,25 – 33	21 - 57
post-aPTT (s)	33,5	30,25 – 38	26 – 160
	Mean	SD+/-	Min.-Maximum
pre-ACT (s)	133,5	12,3	105 – 157
post-ACT (s)	127	10,3	108 – 143

Results

The heparin-protamine-titration assay (Hepcon) showed incomplete heparin reversal in none of the patients (lower detection threshold 0.4U/l). The PPR for aPTT detected 28/36 (77.7%) and the PPR for ACT detected 10/23 (43.5%). The TTC (R-time ratio TEG®6s) detected 33/36 (92%) patients. There was no correlation between the TTC of TEG®6s and the ROTEM® delta ($r=0.089$, $p=0.72$), the PPR of the aPTT ratio ($r=-0.04$, $p=0.81$) or the PPR of the ACT ratio ($r=0.26$, $p=0.21$).