MANUAL OR AUTOMATIC DEFIBRILLATORS FOR THE IN-HOSPITAL SETTING:

SUGGESTED CHOICE CRITERIA

Oey L,^{a,e} Hornix E,^a Beijsterveldt T van,^a Berkom P. van,^b Gullik C van,^c Pronk A,^d Noordergraaf GJ^a

BACKGROUND

Hospitals provide a safe environment, including early defibrillation. Except for high-risk areas, the use of defibrillators is infrequent. Although rapid response teams (RRT) bring their own defibrillator/monitor, the benefit of early shocks strongly suggest immediate availability.¹

The hospital deals with two types of cardiac arrests. Out-patient and visitors represent the OHCA, which tend to be shockable. For in-hospital cardiac arrest (IHCA), the use of automated external defibrillators (AED) is not associated with improved survival, since these are the non-shockable rhythms.² Therefore a distribution of different types of defibrillator is necessary. Table 1 shows the different types of defibrillators and their functionalities.

Currently there are no guidelines as to choice, placement strategies or types/level of defibrillators for in-hospital settings, beside investment costs.²

PURPOSE

Development of rational criteria for type and placement of defibrillators for in-hospital settings.

METHODS

- Literature search concerning in-hospital deployment of defibrillators
- Survey of 72 hospitals in the Netherlands, employees of medical equipment maintenance, resuscitation coordinators and manufacturers using standardized questionnaire on investment and placement strategies
- Interviews with nurses on preferences and criteria

Table 1. Types of defibrillators in the Netherlands

Defibrillator	Type	Functionality	Prize range
AED	Automated or semi-automated	Some have verbal support and print function	€ 1500
Low end	Manual	ECG, SpO ₂ , pacing module	€ 5000
High end	Manual	ETCO ₂ , ECG, SpO ₂ , NIBP, invasive pressures, real-time feedback, registration, data collection	€ 20000
Hybrid	Manual and AED	Same as high or low end defibrillator including AED functionality	€ 10000

RESULTS

- Hospitals: 89% (64) have AED's, low-end and high-end manual defibrillators, 3% (2) have AED's, high-end manual defibrillators, 7% (5) have low-end and high-end defibrillators, and 1% (1) has hybrid defibrillators
- 31% has RRT with high-end manual defibrillator
- Heterogeneity in rational for deployment of defibrillators.
- No validated arguments supporting type/placement of defibrillator (AED vs manual) with wide range of devices-to-patient ratios
- Low incidence of use coupled with low sensitivity to costs
- Guidelines support deployment of a defibrillator within 3 minutes
- Nurses prefer 'real' AED's over manual defibrillators with an AED function

CONCLUSION

Wide range of variation in placement strategies and type of defibrillators for the in-hospital setting.

(Cost) effective placement of defibrillators support use of AED's over manual defibrillators in RRT-based systems, despite their negative effects.²

Users influence type/placement strategies. Rational criteria for suitable and safe placement strategies are listed in Table 2.

Table 2. Criteria for type and placement

Criteria	Type / choice
High risk of VF/pVT (cardio)	Manual + specialized nurses
Unexpected collapse (any cause)	AED if low (defined as < 1 call/month/12 beds)..
User principally BLS trained or ALS trained but inexperienced (i.e. intern/resident)	AED (not manual with AED functionality)
Retrieval time (there and back; based on arrival time RRT ≤ 3 minutes, and compatibility between pads)	AED if more than 3 minutes before arrival RRT. Note RRT expected everywhere within 6 minutes.
In-hospital vs out-of-hospital type patient	AED if out-of-hospital -> outpatient areas
Positioning / crash cart	"easy grab" and all in one required (type device not a factor)

TAKE HOME MESSAGE

Defibrillators are costly, rational criteria for placement are of the essence