A Rapid Response Electrochemical Biosensor for Detecting THC In Saliva

Shalini Prasad, Ph.D.
Department of Bioengineering,
The University of Texas at Dallas,
Richardson, TX 75080, USA.

Background about THC

Methodology

Experimental Results

Marijuana

Δ^9 -tetrahydrocannabinadiol

Cannabidiol

Main
Psychoactive
compound

Long term effects

Impaired
short term
memory

Paranoia

Short term effects

Increased
risk of
chronic
psychosis
disorders

Impaired
motor
coordination

Altered Brain
development

Addiction

THC level in blood drops by 90% within the first hour after smoking, even though the metabolic half-life of THC is estimated to be about four days.

While researchers agree that marijuana impairs driving, but currently there is a dearth of correlation between levels of THC in oral fluid and impairment i.e. **absolute cut-off limits.**

Concentration of THC rapidly decreases to **1-4 ng/mL within 3-4 hours**, therefore there is a requirement for sensors capable of capable of detection at concentrations in the sub ng/mL range

HAIR

semi-intrusive

Determine usage pattern
Dependent on hair length
Storage @ room temperature

THC	Metabolites
3 days → Years	3 days → Years

Verstraete,
Alain G.

SALIVA

least intrusive

THC correlates to impairment
THC & its metabolites present
Originates from oral mucosa depots

THC	Metabolites
Immediate → 34 hours	??

Verstraete,
Alain G.

BLOOD

most intrusive

Most accurate measurement
THC & metabolites present
Standard for DUI investigations

THC	Metabolites
3 minutes → 5 hours	Minutes → 25 Days

Verstraete,
Alain G.

URINE

intrusive

Rapid tests: ↑ false (+)
Only metabolites present
Usage affect results

THC	Metabolites
N/A	Minutes → Days to Months

Verstraete,
Alain G.

Average oral fluid to blood concentration ratio for Δ^9 -THC: 1.2

Threshold limits for Thc in blood/saliva:
0.5-25 ng/mL

- Substance Abuse and Mental Health Services Administration (SAMHSA, Analytes and Their Cutoffs, Federal Register, November 25, 2008 (73 FR 71858), Section 3.4.
- Washington State, Liquor and Cannabis Board- THC limits in blood (DUI).
- Colorado State, Department of Transportation-Cannabis and Driving cut-off limits.
- Drummer OH. Drug testing in oral fluid. *Clin Biochem Rev.* 2006;27(3):147–159.
- D. Gieringer – Feb 2017 – Vsn 2 Feb 25, Cal NORML
- Phillips, J. A. *et al.* (2015) 'Marijuana in the Workplace: Guidance for Occupational Health Professionals and Employers: Joint Guidance Statement of the American Association of Occupational Health Nurses and the American College of Occupational and Environmental Medicine', *Workplace Health & Safety*, 63(4), pp. 139–164. doi: 10.1177/2165079915581983.

A) Sensor Schematic

Electrode Geometry

B)

C)

D)

Electrochemical
Impedance
spectroscopy
(EIS)

Complex
Impedance
Spectrum

EIS

- Frequency based probing
- Captures perturbation at **electrode-electrolyte interface**
- Quantification of biomolecular binding effects occurring at the electrode interface as a function of **impedance modulation**.

Complex Impedance Calculation:

$$Z = \frac{V_{in}}{I_{out}} = \frac{V \sin(\omega t)}{I \sin(\omega t + \phi)}$$

Broad Takeaways:

- Dose dependent impedance response observed across all sample groups (PBS, Synthetic saliva (pH 4,6), Human saliva).
- For human saliva- **Linear dynamic range: 100pg/mL to 100ng/mL**
- Limit of Detection (LoD): **100pg/mL**
- Within the dynamic range, the impedance changes are **significantly distinguishable** from each other.
- **4 ng/mL** which is set as the oral fluid drug screening **cut-off limit as per SAMSHA guidelines** can be effectively screened using this platform.

Objective

Train a model to categorize our samples as THC (+) or THC (-)

Insight

- Supervised Learning implies we are given **known** samples that we then label to train our classifier
- Classification algorithms divide data into subsets in order to verify model accuracy
 - Training set**
 - Validation (Test) set**
- Feature Engineering and pre-processing of data can improve model accuracy and speed

Model Building

In the model building stage we feed the algorithm our input frequency spectrums and our **training set** with labels: THC (+) = 1 & THC (-) = 0

The model then looks for the best way to split the inputs between our samples labelled '1' & '0'

Model Testing

In the model testing stage we deploy the algorithm on 'unseen' data from the **validation set** to predict the labels, then we compare to the actual result

Binary Classification Model Performance

- **Accuracy:** How often is the classifier correct?

$$\frac{(TP+TN)}{total}$$

- **Misclassification Rate:** How often is the classifier wrong?

$$\frac{(FP+FN)}{total}$$

(TP-True positive, TN-True negative, FP-False positive, FN-False negative)

Confusion Matrix

		Actual Value	
		THC (+)	THC (-)
Predicted Value	total		
	THC (+)	True Positives	False Positives
	THC (-)	False Negatives	True Negatives

Receiver Operator Characteristics

Optimized feature and frequency selection

- Based on fast correlation based filter (FCBF), change in imaginary impedance (dz'') with respect to the antibody measurement at 9.9 Hz was selected as the best feature for separating the data by class (THC+/-)

(b)

Classifier Name	AUC	Threshold	TPR	FPR
Logistic Regression	0.911	0.39	97.1	60.0
		0.722	88.6	15.0
Logistic Regression with K-folds cross-validation	0.916	0.303	97.1	60.0
		0.769	88.6	10.0
Linear kernel SVM	0.911	0.214	97.1	60.0
		0.829	88.6	15.0
Radial Bias Function kernel SVM	0.951	0.621	97.1	15.0
		0.797	94.3	5.0

- This work utilizes non-faradaic EIS to rapidly and precisely detect the THC-BSA hapten spiked in human saliva with high specificity and sensitivity. (LoD: 100 ng/mL)
- A portable, battery-powered device interfaced with a one-time use saliva swab embedded with the THC detection sensor for easier and faster (< 3-5 minutes) roadside testing when implemented using low-power EIS hardware.

- Implementing a non-linear RBF kernel SVM algorithm to perform binary classification demonstrated a very high general performance (AUC = 0.951) at predicting the presence of THC within human saliva.

- Volkow ND, Baler RD, Compton WM, Weiss SR. Adverse health effects of marijuana use. *N Engl J Med*. 2014;370(23):2219–2227. doi:10.1056/NEJMra1402309
- Stevenson, H., Bacon, A., Joseph, K.M. *et al*. A Rapid Response Electrochemical Biosensor for Detecting Thc In Saliva. *Sci Rep* **9**, 12701 (2019). <https://doi.org/10.1038/s41598-019-49185-y>
- Hall, W. & Degenhardt, L. Adverse health effects of non-medical cannabis use. *Lancet (London, England)* **374**, 1383–1391 (2009).
- Hartman, R. L. & Huestis, M. A. Cannabis Effects on Driving Skills. *Clin. Chem.* 59, <https://doi.org/10.1373/clinchem.2012.194381> (2013).
- Statistics, H. 2014 - SAMHSA - NSDUH 2013 results. doi:NSDUH Series H-41, HHS Publication No. (SMA) 11–4658. (2013).
- Substance Abuse and Mental Health Services Administration (SAMHSA, Analytes and Their Cutoffs, Federal Register, November 25, 2008 (73 FR 71858), Section 3.4.
- Washington State, Liquor and Cannabis Board- THC limits in blood (DUI).
- Colorado State, Department of Transportation-Cannabis and Driving cut-off limits.
- Drummer OH. Drug testing in oral fluid. *Clin Biochem Rev.* 2006;27(3):147–159.
- D. Gieringer – Feb 2017 – Vsn 2 Feb 25, Cal NORML
- Phillips, J. A. et al. (2015) ‘Marijuana in the Workplace: Guidance for Occupational Health Professionals and Employers: Joint Guidance Statement of the American Association of Occupational Health Nurses and the American College of Occupational and Environmental Medicine’, *Workplace Health & Safety*, 63(4), pp. 139–164. doi: 10.1177/2165079915581983.
- Sharma P, Murthy P, Bharath MM. Chemistry, metabolism, and toxicology of cannabis: clinical implications. *Iran J Psychiatry*. 2012;7(4):149–156.